
# D LUDLUM MEASUREMENTS, INC.

## Model 194 Dose Equivalent Rate Meter

### Features

- Accurate (Energy-Flat) Ambient Dose Equivalent Rate Measurements at Background Levels
- Eliminates High False Readings of Typical microR Meters
- Rugged Scintillator, Will Not Leak
- Low Range: Background to 19.99 μSv/h (1999 μrem/hr)
- High Range: 1 to 1999 µSv/h (0.1 to 199.9 mrem/hr)
- Improved Replacement for Bicron Microrem
- "/E" Extended Version Has Low-Energy, 3.2 mg/cm<sup>2</sup>
  Window for Operation Below 50 keV



#### **Specifications**

Model 194 (μrem/hr) PN: 48-4364 Model 194/E (μrem/hr) (Extended Detector) PN: 48-4388

INDICATED USE: Low-level surveys

DETECTOR: Internal rugged plastic scintillator, EJ-256

**ENERGY RESPONSE (Model 194 only):** Energy-flat within  $\pm$  30% to ambient dose equivalent from 50 to 1300 keV

LOW RANGE: Background to 19.99 µSv/h (1999 µrem/hr)

**HIGH RANGE:** 1 to 1999 μSv/h (0.1 to 199.9 mrem/hr)

DISPLAY: 3½-digit LCD with 10 mm (0.4 in.) digits, backlight, and protective bezel

#### **CONTROLS:**

- Rotary Selector: switches between OFF, battery check, low range, high range
- **Response Switch:** toggles between FAST (3 seconds) and SLOW (15 seconds)
- Audio Switch: audio ON/OFF, built-in unimorph speaker. Sound frequency increases as meter reading increases.
- Lamp Switch: display backlight ON/OFF
- **Calibration Controls:** accessible from front of instrument (protective cover provided)
- **OVERRANGE PROTECTION:** When reading is out of the display range, display shows overrange (leftmost "1" and no other digits)

**POWER:** Two each "AA" cell batteries (externally accessible from the front of the instrument)

**BATTERY LIFE:** Approximately 200 hours, green BAT OK light indicates adequate battery power

TEMPERATURE RANGE: -10 to 50 °C (14 to 122 °F)

ENVIRONMENTAL RATING: Water-resistant, IP rating of 65


**CONSTRUCTION**: Drawn-and-cast aluminum case with beige powdercoating

**SIZE**: 19.0 x 8.9 x 21.6 cm (7.5 x 3.5 x 8.5 in.) **WEIGHT**: 1.1 kg (2.5 lb), including batteries

Options

Headphone Modification Part Number: 4475-118

Model 194 Energy Response Compared to Typical Micro R Meter 1000% Xe< -20% Relative to 662 100% Model 194 Response Typical Micro R Detecto 10% 1% 10 100 1000 KeV



Shoulder Strap Part Number: 4475-120

Ludium Measurements, Inc. P.O. Box 810, Sweetwater, Texas 79556

Web: http://www.ludlums.com Tel: 800-622-0828 / 325-235-5494 / Fax: 325-235-4672 / Email: sales@ludlums.com Note: specifications subject to change without notification. We are not responsible for errors or omissions.

**Model 194/S (μSv/h)** PN: 48-4364-1 **Model 194/SE (μSv/h) (Extended Detector)** PN: 48-4388-1